



Overview of Ambu's Training Devices in relation to content of Latex and REACH Regulated SVHC substances.

According to the European Chemical regulation (REACH) Ambu is obligated to inform about products containing listed SVHC substances (Substances of Very High Concern) in concentrations above 0.1% (w/w). See the following website for the current version of the SVHC list:

<http://echa.europa.eu/web/guest/candidate-list-table>

The table below provides an overview of all Ambu Training devices in relation to content of REACH Regulated SVHC substances in concentrations above 0.1% (w/w). The table also includes information on which products are made with Natural Rubber Latex. Ambu products presented in this list will be labeled as indicated below. All products that has “No” in the SVHC column (column 2) does not contain any substances currently on the SVHC list.

Product	Contents of SVHC above 0.1% (w/w) and description of substance	Natural Rubber Latex	Symbol on label
Emergency Care and Training			
BLS:			
Ambu® Man	No	No	
Ambu® Baby Manikin	No	No	
Ambu® Defib Trainer System	No	No	
Ambu® CPR Pal	No	No	
Ambu® Man Compression	No	No	
Ambu® Man Basic	No	No	
Ambu® SAM	No	No	
Ambu® Junior	No	No	
Ambu® Man Wireless - Next Generation	No	No	
Ambu® Man Instrument	No	No	
Ambu® Man Defib – Next Generation	No	No	
Ambu® Man School	No	No	
Ambu® Patented Hygienic System	No	No	
Ambu® CPR software	No	No	
Ambu® AmbuMan Basic iQF	No	No	
Ambu® Baby iQF	No	No	
AmbuMan® SAM iQF	No	No	
AmbuMan® Compression iQF	No	No	
ALS:			
Ambu® Airway Man	No	No	
Ambu® Airway Management Trainer	No	No	
Ambu® IV Trainer	No	Yes	
Ambu® Kit i.o. Access	No	No	
Ambu® Man Airway Wireless – Next Generation	No	No	
Ambu® Man Airway Instrument	No	No	
Ambu® Man Advanced IV – Next Generation	No	Yes	
Ambu® Man Advanced – Next Generation	No	No	
Ambu® Cardiac Care Trainer	No	Yes	

